EVA 450 and 600 Mobile Lift
Competency Assessment

Competency Statement: After attending this orientation, the clinician will be able to properly use specialty equipment according to manufacturer's recommendations. If the performance criteria are not met, additional training will be provided. If the caregiver has any questions always consult the product owner manual prior to use or contact your Handicare Representative.

Directions: Check the appropriate box for each performance criteria. Schedule retraining as indicated.

Describe and demonstrate equipment features	**Verbalized understanding**	**Performed**	**Needs review**
Safety checks Mast and boom secure All nuts and bolts secure Sling free of defects, tears or fraying of loops			
Operation of EVA 450 and 600 Mobile Lift			
1. Identify lift capacity is within limits for patient			
2. Position lift around bed or chair using leg spreader button as			
3. Sling selection and application competency • Highback Sling • Limb Sling • Repostiioning			
4. Select appropriate sling and position under patient			
5. Attach sling to sling bar* Use the same colored loop on the patient's left side as the patient's right side. Select shorter shoulder loops and longer leg loops for upright seating. Select longer shoulder loops and shorter leg loops for more reclined seating.			
6. Ensure wheels are UNLOCKED (Explain rationale)			
7. Press "up" button on lift or hand control			
8. Move lift to receiving surface and center patient			
9. Press "down" button on lift or hand control			
10. Unhook sling from sling bar (verbalize method)			
11. Battery: Locate and remove, operate red on/off switch, verbalize understanding of charging			
12. Locate and operate emergency down			
13. Observe and verbalize understanding of Lift/Transfer from chair, bed, floor			
14. Demonstrate proper body mechanics when moving lift with patient			
15. Caregiver can verbalize as to where the user manual is kept on the unit			

*If attaching a seated sling or limb sling, the sling bar will be parallel with the patient's shoulders. Repositioning sling types should be connected with the sling bar perpendicular to the patient's shoulders.

I understand how to use the equipment and all of my questions have been answered. If I have any questions prior to use of lift I will consult the owner/user manual for the EVA floor lift product.

Name: _____ Title_____ Signature_____

Evaluator Signature: _____Date:_____

SystemRoMedic